


EU Declaration of Conformity
EU-Konformitätserklärung
Déclaration de Conformité UE



Electromach b.v. • Jan Tinbergenstraat 193 • 7559 SP Hengelo, The Netherlands
 declares in its sole responsibility, *erklärt in alleiniger Verantwortung, déclare sous sa seule responsabilité,*

that the product: <i>dass das Produkt:</i> <i>que le produit:</i>	Mass Flow meter/Controller, mini CORI-FLOW
Type(s), typ(en), type(s):	XM1.-...-...
Order Number: <i>Auftragsnummer:</i> <i>numéro de commande</i>	N.A.
is in conformity with the requirements of the following directives and standards. <i>mit den Anforderungen der folgenden Richtlinien und Normen übereinstimmt.</i> <i>est conforme aux exigences des directives et des normes suivantes.</i>	
Directive(s) <i>Richtlinie(n)</i> <i>Directive(s)</i>	Standard(s) <i>Norm(en)</i> <i>Norme(s)</i>
2014/34/EU: ATEX 114 Directive <i>2014/34/EU: ATEX 114 Richtlinie</i> <i>2014/34/EU: Directive 114 ATEX</i>	EN 60079-0:2012/A11:2013 EN 60079-1: 2014 EN 60079-7: 2007 EN 60079-11: 2012 EN 60079-18: 2015
Marking, Kennzeichnung, marquage:	II 2 G Ex d e IIB T6 Gb (Standard version and Extended version without IS display and isolating repeater)  II 2(1) G Ex d e mb [ia Ga] IIB T4 Gb CE 0158 (Extended version without IS display) II 2(1) G Ex d e ia mb [ia Ga] IIB T4 Gb (Extended version with IS display)
EC Type Examination Certificate: <i>EG-Baumusterprüfbescheinigung:</i> <i>Attestation d'examen CE de type:</i>	DEKRA 12ATEX0144 X (DEKRA Certification B.V., Meander 1051, 6825 MJ Arnhem, The Netherlands)
Ambient temperature, <i>Umgebungstemperatur,</i> <i>Température ambiante</i>	Standard Version 0°C...+55°C (standard ambient) -20°C...+55°C (low ambient) Extended version 0°C...+40°C (standard ambient) -20°C...+40°C (low ambient)
2014/30/EU: EMC Directive <i>2014/30/EU: EMV-Richtlinie</i> <i>2014/30/CE: Directive CEM</i>	EN 61439-1 EN 61439-2

Hengelo, 05-04-2017

Place and date
Ort und Datum
Lieu et date

C.L.L. Cameron
Quality manager
Leiter Qualitätssicherung
Directeur Assurance de Qualité